

LAW OFFICES OF
FAY KAPLUN & MARCIN, LLP
INTELLECTUAL PROPERTY LAW

150 BROADWAY, SUITE 702
NEW YORK, NEW YORK 10038
PHONE: (212) 619-6000
FAX: (212) 619-0276
WWW.FKMIPLAW.COM

RECEIVED
CENTRAL FAX CENTER

JUN 01 2007

FACSIMILE COVER SHEET

FAX NO : (571) 273-8300
TO : Commissioner for Patents
Mail Stop: Appeal Brief-Patents
FROM : Oleg F. Kaplun, Esq. of Fay Kaplun & Marcin, LLP
DATE : June 1, 2007
SUBJECT : U.S. Patent Appln. Serial No. 10/762,715
for *Valved Catheter to Bypass Connector*
Our Ref.: 10123/03601

NUMBER OF PAGES INCLUDING COVER : 19

MESSAGE:

Please see attached.

Thank you.

IF ANY PAGES WERE NOT RECEIVED OR ARE ILLEGIBLE, PLEASE CALL (212) 619-6000 AS SOON AS POSSIBLE

The information contained in this facsimile message is attorney privileged and confidential information intended only for the use of the individual or entity named above. If the reader of this message is not the intended recipient or the employee or agent responsible to deliver it to the intended recipient, you are hereby notified that any dissemination, distribution or copying of this communication is strictly prohibited. If you have received this communication in error, please notify us by telephone, and return the original message to us at the above address via the U.S. Postal Service. We will reimburse any costs you incur in notifying us and returning the message to us. Thank you.

Attorney Docket No. 10123/03601 (03-225)


IN THE UNITED STATES PATENT AND TRADEMARK OFFICE

Applicant : DiMatteo et al.
Serial No. : 10/762,715
Filing Date : January 22, 2004
For : Valved Catheter to Bypass Connector
Art Unit : 3767
Confirmation No. : 5203
Examiner : Phillip A. Gray

RECEIVED
CENTRAL FAX CENTER

JUN 01 2007

Mail Stop: Appeal Brief - Patent
Commissioner for Patents
P.O. Box 1450
Alexandria, VA 22313-1450


Certificate of Facsimile/First Class Mail	
I hereby certify that this correspondence is being transmitted via facsimile or First Class Mail addressed to:	
Mail Stop: Appeal Brief-Patent Commissioner for Patents P.O. Box 1450 Alexandria, VA 22313-1450 (671) 273-8300	
By:  Oleg F. Kaplun Reg. No. 45,559	Date: June 1, 2007

TRANSMITTAL

In response to the Notification of Non-Compliant Appeal Brief mailed on May 9, 2007, transmitted herewith please find a revised Appeal Brief for filing in the above-identified application. No fees are believed to be required. However, the Commissioner is hereby authorized to charge the **Deposit Account of Fay Kaplun & Marcin, LLP NO. 50-1492** for any additional required fees. A copy of this paper is enclosed for that purpose.

Respectfully submitted,

Dated: June 1, 2007

By: 
Oleg F. Kaplun, Reg. 45,559

Fay Kaplun & Marcin, LLP
150 Broadway, Suite 702
New York, NY 10038
Tel: (212) 619-6000
Fax: (212) 619-0276

**RECEIVED
CENTRAL FAX CENTER****JUN 01 2007****PATENT**
Attorney Docket No.: 10123/03601**IN THE UNITED STATES PATENT AND TRADEMARK OFFICE
BEFORE THE BOARD OF PATENT APPEALS AND INTERFERENCES**

In re Application of:)	
)	
DiMatteo et al.)	
)	
Serial No.: 10/762,715)	Group Art Unit: 3767
)	
Filed: January 22, 2004)	Examiner: Phillip Gray
)	
For: VALVED CATHETER TO)	Board of Patent Appeals and
BYPASS CONNECTOR)	Interferences
)	

Mail Stop: Appeal Brief - Patents
Commissioner for Patents
P.O. Box 1450
Alexandria, VA 22313-1450

APPEAL BRIEF UNDER 37 C.F.R. § 41.37

In support of the Notice of Appeal filed January 9, 2007, and pursuant to 37 C.F.R. § 41.37, Appellants present their appeal brief in the above-captioned application.

This is an appeal to the Board of Patent Appeals and Interferences from the Examiner's final rejection of claims 1-5, 7-19, and 21-24 in the final Office Action dated September 29, 2006. The appealed claims are set forth in the attached Claims Appendix.

Serial No.: 10/762,715
Group Art Unit: 3767
Attorney Docket No.: 10123/03601

1. Real Party in Interest

This application is assigned to Boston Scientific Scimed, Inc., the real party in interest.

2. Related Appeals and Interferences

There are no other appeals or interferences which would directly affect, be directly affected by, or have a bearing on the instant appeal.

3. Status of the Claims

Claims 1-5, 7-19, and 21-24 stand rejected in the Final Office Action. The final rejection of claims 1-5, 7-19, and 21-24 is being appealed.

4. Status of Amendments

All amendments submitted by the Appellants have been entered.

5. Summary of Claimed Subject Matter

The present invention describes, as recited in independent claim 1, a connector for injecting fluid to a catheter. The connector (100) comprises an attachment portion (104) adapted to fluidly couple to a source of pressurized fluid. (See Specification, p. 6, ll. 14-28; Fig. 1). The connector comprises a bypass element (e.g., 108) fluidly connected to the attachment portion. (See *Id.*, p. 7, l. 1 - p. 8, l. 3; Figs. 1-2). The bypass element is adapted to open a valve of the catheter to permit fluid to flow into the catheter without impinging on the valve. (See *Id.*). The

Serial No.: 10/762,715
Group Art Unit: 3767
Attorney Docket No.: 10123/03601

connector comprises an overpressure control element (e.g., 306) adapted to maintain a pressure of fluid within the connector below a predetermined threshold level. (See Id., p. 9, ll. 7-16; Fig. 1).

The present invention describes, in an independent claim 18, a fluid coupler. The coupler (100) comprises an elongated tube (102) extending between a first end (104) adapted for fluid connection to a power injector and a second end adapted for fluid connection (via extensions 106) to a catheter (200) including a valve (202) in a proximal part thereof. (See Id., p. 6, ll. 14-28; p. 7, l. 22 - p. 8, l. 3; Figs. 1-2). The second end is insertable into the catheter beyond the valve thereof so that fluid passes through the fluid coupler into the catheter to a distal end thereof without passing through the valve. (See Id.). The coupler comprises a pressure control element (306) adapted to limit a fluid pressure within the coupler to a predetermined threshold level. (See Id., p. 9, ll. 7-16; Fig. 1).

6. Grounds of Rejection to be Reviewed on Appeal

- I. Whether claims 1-5, 7-19, and 21-24 are unpatentable under 35 U.S.C. § 103(a) over U.S. Pat. No. 4,142,525 to Binard et al. ("Binard") on view of U.S. Pat. No. 6,227,200 to Crump et al. ("Crump").
- II. Whether claims 14-17 are unpatentable under 35 U.S.C. § 103(a) over U.S. Pat. No. 4,142,525 to Binard et al. ("Binard") on view of U.S. Pat. No. 6,227,200 to Crump et al. ("Crump").

Serial No.: 10/762,715
Group Art Unit: 3767
Attorney Docket No.: 10123/03601

7. Argument

I. The Rejection of Claims 1-5, 7-19, and 21-24 Under 35 U.S.C. § 103(a) as Unpatentable Over U.S. Pat. No. 4,142,525 to Binard et al. In view of U.S. Pat. No. 6,227,200 to Crump et al. Should Be Reversed

A. The Examiner's Rejection

In the Final Office Action, claims 1-5, 7-19, and 21-24 were rejected under 35 U.S.C. 103(a) as unpatentable over Binard in view of Crump. (See 9/29/06 Office Action, p. 3, ll. 4-6).

Binard describes a syringe assembly that comprises a chamber for retaining fluid, a plunger to pump fluid out of the chamber, and a means for limiting the amount of pressure generated by the syringe during pumping of the fluid. (See Binard, Abstract). Binard includes an extension that has a passageway that communicates with the syringe chamber and the hollow needle. (See *Id.*, col. 3, ll. 26-28). The passageway has a flexible balloon that communicates with the passageway so that it inflates when a predetermined pressure has been reached in the passageway or chamber of the syringe. (See *Id.*, col. 3, ll. 40-44). The balloon serves to limit the amount of pressure generated by the syringe, and prevents ejection of further fluid in the needle and sub-dural space after the predetermined pressure has been attained. (See *Id.*, col. 4, ll. 6-10). In other words, Binard discloses an overpressure control element that alleviates increases in pressure of fluid by determining if the pressure of the passageway has risen *above* a threshold level.

Serial No.: 10/762,715
Group Art Unit: 3767
Attorney Docket No.: 10123/03601

Crump describes a device used for suction. The respiratory suction apparatus catheter includes a manifold for attachment to the distal hub of an endotracheal tube to form a ventilation circuit, a catheter tube which is displaceable through the manifold and into the endotracheal tube to suction secretions from the tube and lungs, and a valve mechanism disposed adjacent the ventilation circuit to minimize the draw of air from the ventilation circuit of a patient while the catheter is being cleaned. (See Crump, abstract). Specifically, Crump discloses a “respiratory suction catheter apparatus [that] includes a manifold 404 and a catheter 408 which is movable through the manifold to suction secretions from a patient’s lungs.” (See Crump, col. 11, lines 8-11). Thus, a negative pressure is generated in order to operate the device in Crump.

B. Binard in View of Crump Does Not Disclose or Suggest a Bypass Element Adapted to Open a Valve of the Catheter to Permit Fluid to Flow into the Catheter without Impinging on the Valve, as Recited in Claim 1.

Claim 1 recites a connector for injecting fluid to a catheter, comprising a “bypass element fluidly connected to the attachment portion, the bypass element being adapted to open a valve of the catheter to permit fluid to flow into the catheter without impinging on the valve” in combination with “an overpressure control element adapted to maintain a pressure of fluid within the connector below a predetermined threshold level.”

In the final rejection, the Examiner correctly stated that Binard does not disclose the valve type that permits fluid to flow into the catheter without impinging the valve. (See 9/29/06 Office Action, p. 4, ll. 8-9). Specifically, Binard does not disclose a “bypass element

Serial No.: 10/762,715
Group Art Unit: 3767
Attorney Docket No.: 10123/03601

adapted to open a *valve of the catheter* to permit fluid to flow into the catheter without impinging on the valve,” as recited in claim 1. The Examiner attempted to cure this deficiency with Crump. However, it is respectfully submitted that because the valve disclosed in Crump is not part of the catheter but exists as a separate unit, the Examiner’s claim is unfounded.

Appellants respectfully submit that Crump is directed to a mechanism for cleaning the tip of a catheter and that the catheter of Crump includes no valve. (See Crump, col. 1, ll. 12-20). Specifically, the valve 424 cited by the Examiner is located in a manifold which is received around the catheter 408 and no valve is located in the catheter 408. (See Id., col. 11, ll. 18-24; Figs. 5A-C). The manifold 404 shown in Figs. 5A-C includes a port 412d which receives a connector or adaptor 420 into which the catheter 408 is inserted. The valve 424 extends across the opening of the adaptor 420 and is forced open as the catheter 408 is inserted therethrough and which closes when the catheter 408 is withdrawn. (See Id., col. 11, ll. 11-23). Crump explicitly states that the “valve 424 is attached to the adaptor 420 by a flexible base 428,” not that the valve is located in or attached to the catheter 408 in any way. (See Id., col. 11, ll. 23-24).

The Examiner asserted that Binard in view of Crump teaches a valve of a catheter which is opened to permit fluid to flow into the catheter. (See 12/4/06 Advisory Action, p. 2, ll. 15-17). Specifically, the Examiner stated that this assertion was based on the premise that a claim limitation is given the broadest interpretation during examination. However, Appellants respectfully submit that even given the broadest interpretation, Crump does not teach a “*valve of the catheter* is opened to permit fluid to flow into the catheter,” as recited in claim 1. As

Serial No.: 10/762,715

Group Art Unit: 3767

Attorney Docket No.: 10123/03601

discussed above, the valve 424 of Crump is part of the manifold. The valve 424 is described with great detail in that it creates a seal with an outer circumference of the catheter 408. The created seal is a feature of the valve since a minimized space between the valve and the catheter is desired in Crump. (See Crump, col. 8, ll. 3-5; col. 11, ll. 35-38). Because Crump is directed toward cleaning the catheter prior to removal, Crump focuses on the outer circumference of the catheter 408. Furthermore, there is no disclosure in the entirety of Crump that a valve is located in or part of the catheter 408. A valve located in the catheter 408 defeats one of the features of Crump. Specifically, Crump utilizes *lateral* apertures 248 on the distal end of the catheter 408. Those skilled in the art will understand that a valve located in the catheter 408 of Crump would serve absolutely no purpose and would not benefit the respiratory suction catheter apparatus. Therefore, it is respectfully submitted that Binard and Crump, taken either alone or in combination, do not disclose or suggest a bypass element “adapted to open a *valve of the catheter* to permit fluid to flow into the catheter without impinging on the valve,” as recited in claim 1.

Thus, Appellants respectfully request that the Board overturn the Examiner’s rejection under 35 U.S.C. § 103(a) of claim 1 and claims 2-5 and 7-17 which depend directly or indirectly therefrom.

Claim 18 also includes limitations that distinguish over Binard in view of Crump in a manner substantially similar to the reasons stated above in regard to claim 1. Specifically, claim 18 recites “an elongated tube extending between a first end adapted for fluid connection to a power injector and a second end adapted for fluid connection to *a catheter including a valve in*

Serial No.: 10/762,715
Group Art Unit: 3767
Attorney Docket No.: 10123/03601

a proximal part thereof, the second end being insertable into the catheter beyond the valve thereof so that fluid passes through the fluid coupler into the catheter to a distal end thereof without passing through the valve." Thus, claim 18 recites a device for insertion into a catheter to open a valve in the catheter and it is respectfully submitted that neither Binard nor Crump either shows or suggests such an arrangement.

Thus, Appellants respectfully request that the Board overturn the Examiner's rejection under 35 U.S.C. § 103(a) of claim 18 and claims 19 and 21-24 which depend therefrom.

II. The Rejection of Claims 14-17 Under 35 U.S.C. § 103(a) as Unpatentable Over U.S. Pat. No. 4,142,525 to Binard et al. In view of U.S. Pat. No. 6,227,200 to Crump et al. Should Be Reversed

A. The Examiner's Rejection

In the Final Office Action, claims 14-17 were rejected under 35 U.S.C. 103(a) as unpatentable over Binard in view of Crump. (See 9/29/06 Office Action, p. 4, ll. 21-22). Binard and Crump were discussed above.

B. Binard in View of Crump Does Not Disclose or Suggest a Bypass Element Adapted to Open a Valve of the Catheter to Permit Fluid to Flow into the Catheter without Impinging on the Valve, as Recited in Claim 1.

In the final rejection, the Examiner correctly stated that Binard in view of Crump does not disclose or suggest the threshold pressure levels between approximately 300 p.s.i. to 40

Serial No.: 10/762,715

Group Art Unit: 3767

Attorney Docket No.: 10123/03601

p.s.i. (See 9/29/06 Office Action, p. 4, l. 22 - p. 5, l. 2). The Examiner asserts that it would have been obvious to one having ordinary skill in the art to use these threshold pressure levels. The Examiner had already rejected claims 14-17 under Binard in view of Crump but rejects these claims again stating this obviousness argument. However, it is respectfully submitted that the Examiner's obviousness assertion is moot. Claims 14-17 indirectly depend from claim 1. Therefore, claims 14-17 includes all of the elements and limitations of its respective independent claim. As discussed above, Binard and Crump, taken either alone or in combination, do not disclose or suggest a bypass element "adapted to open a *valve of the catheter* to permit fluid to flow into the catheter without impinging on the valve," as recited in claim 1. Thus, it is respectfully submitted that claims 14-17, which depend indirectly from claim 1, are allowable.

RECEIVED
CENTRAL FAX CENTER

JUN 01 2007


Serial No.: 10/762,715
Group Art Unit: 3767
Attorney Docket No.: 10123/03601

8. Conclusions

For the reasons set forth above, Appellants respectfully request that the Board reverse the final rejections of the claims by the Examiner under 35 U.S.C. § 103(a) and indicate that claims 1-5, 7-19, and 21-24 are allowable.

Respectfully submitted,

Date: June 1, 2007


By: _____
Oleg F. Kaplun (Reg. No. 45,559)

Fay Kaplun & Marcin, LLP
150 Broadway, Suite 702
New York, NY 10038
Tel: (212) 619-6000
Fax: (212) 619-0276

RECEIVED
CENTRAL FAX CENTER

JUN 01 2007

Serial No.: 10/762,715

Group Art Unit: 3767

Attorney Docket No.: 10123/03601

CLAIMS APPENDIX

1. (Rejected) A connector for injecting fluid to a catheter, comprising:

an attachment portion adapted to fluidly couple to a source of pressurized fluid;

a bypass element fluidly connected to the attachment portion, the bypass element being adapted to open a valve of the catheter to permit fluid to flow into the catheter without impinging on the valve; and

an overpressure control element adapted to maintain a pressure of fluid within the connector below a predetermined threshold level.
2. (Rejected) The connector according to claim 1, wherein the bypass element comprises an elongated tubular component insertable into the catheter through the valve of the catheter.
3. (Rejected) The connector according to claim 2, wherein the elongated tubular component has a diameter selected to fit in a flow opening of the valve of the catheter.
4. (Rejected) The connector according to claim 2, wherein the elongated tubular component is hypotube.
5. (Rejected) The connector according to claim 2, wherein the elongated tubular component includes an outlet which, when the elongated tubular component is inserted into the catheter through the valve, is located distally of the valve.

Serial No.: 10/762,715

Group Art Unit: 3767

Attorney Docket No.: 10123/03601

7. (Rejected) The connector according to claim 1, wherein the overpressure control element comprises a pressure relief valve.
8. (Rejected) The connector according to claim 1, wherein the overpressure control element comprises a controlled failure element designed to fail when a fluid pressure therein reaches the threshold level.
9. (Rejected) The connector according to claim 8, wherein the controlled failure element is an extension tube.
10. (Rejected) The connector according to claim 1, further comprising an external collection jacket disposed around the overpressure control element.
11. (Rejected) The connector according to claim 1, wherein the bypass element is adapted to open a pressure actuated safety valve of a venous catheter.
12. (Rejected) The connector according to claim 1, wherein the attachment portion is adapted to connect to a contrast media power injection system.
13. (Rejected) The connector according to claim 1, wherein the threshold level is selected to be less than a burst pressure of a catheter with which the connector is to be used.
14. (Rejected) The connector according to claim 13, wherein the threshold level is approximately 300 psi.
15. (Rejected) The connector according to claim 14, wherein the threshold level is approximately 100 psi.

Serial No.: 10/762,715

Group Art Unit: 3767

Attorney Docket No.: 10123/03601

16. (Rejected) The connector according to claim 13, wherein the threshold level is approximately 80 psi.

17. (Rejected) The connector according to claim 16, wherein the threshold level is approximately 40 psi.

18. (Rejected) A fluid coupler comprising:

an elongated tube extending between a first end adapted for fluid connection to a power injector and a second end adapted for fluid connection to a catheter including a valve in a proximal part thereof, the second end being insertable into the catheter beyond the valve thereof so that fluid passes through the fluid coupler into the catheter to a distal end thereof without passing through the valve; and

a pressure control element adapted to limit a fluid pressure within the coupler to a predetermined threshold level.

19. (Rejected) The coupler according to claim 18, wherein the elongated tube is a hypotube.

21. (Rejected) The coupler according to claim 18, wherein the pressure control element comprises a section having a burst pressure lower than a burst pressure of the catheter.

22. (Rejected) The coupler according to claim 18, wherein the pressure control element comprises an extension tube connected to the first end.

23. (Rejected) The coupler according to claim 18, wherein the pressure control element comprises a pressure relief valve.

Serial No.: 10/762,715

Group Art Unit: 3767

Attorney Docket No.: 10123/03601

24. (Rejected) The coupler according to claim 18, further comprising a fluid collection jacket surrounding the pressure control element.

Serial No.: 10/762,715
Group Art Unit: 3767
Attorney Docket No.: 10123/03601

EVIDENCE APPENDIX

No evidence has been entered or relied upon in the present appeal.

Serial No.: 10/762,715

Group Art Unit: 3767

Attorney Docket No.: 10123/03601

RELATED PROCEEDING APPENDIX

No decisions have been rendered regarding the present appeal or any proceedings related thereto.